For Research Use Only

## NGAL Recombinant antibody, PBS Only (Capture)



Catalog Number:82841-9-PBS

## Catalog Number: GenBank Accession Number: **Purification Method: Basic Information** 82841-9-PBS NP\_005555.2 Protein A purfication Size: GeneID (NCBI): CloneNo.: 1mg/ml 3934 2N10 Source: UNIPROT ID: Rabbit P80188 Full Name: Isotype: lgG lipocalin 2 Calculated MW: 23kd **Applications Tested Applications:** Sandwich ELISA, Indirect ELISA, Sample test

Sandwich ELISA, Indirect ELISA, Sample Species Specificity: human

## **Background Information**


Storage

Storage: Store at -80°C. The product is shipped with ice packs. Upon receipt, store it immediately at -80°C Storage Buffer: PBS Only


For technical support and original validation data for this product please contact:T: 4006900926E: Proteintech-CN@ptglab.comW: ptgcn.com

This product is exclusively available under Proteintech Group brand and is not available to purchase from any other manufacturer.


## Selected Validation Data




Sandwich ELISA standard curve of MP00770-1, NGAL Recombinant Matched Antibody Pair - PBS only. 82841-9-PBS was coated to a plate as the capture antibody and incubated with serial dilutions of standard Eg0268. 82841-2-PBS was HRP conjugated as the detection antibody. Range: 19.5-1250 pg/mL




Human peripheral blood mononuclear cells (PBMC) were cultured unstimulated or stimulated with 10  $\mu$  g/mL PHA for 1 days. The mean NGAL concentration was determined to be 18,080.8 pg/mL in unstimulated PBMC supernatant, 32,268.8 pg/mL in PHA stimulated PBMC supernatant.



Saliva of eight individual healthy human donors was measured. The NGAL concentration of detected samples was determined to be 158,732.5 pg/mL with a range of 52,136.9 - 388,400.2 pg/mL





Urine of eight individual healthy human donors was measured. The NGAL concentration of detected samples was determined to be 11,654.6 pg/mL with a range of 1,964.2 - 21,150.4 pg/mL

Plasma of fifteen individual healthy human donors was measured. The NGAL concentration of detected samples was determined to be 38,260.8 pg/mL with a range of 17,778.7 - 90,593.8 pg/mL